

LARGE RANGE BACK PLUNGER TYPE DIAL TEST INDICATOR



**INSPECTION
CERTIFICATE**
TRACEABLE TO NIST

JEWELLED
BEARING




2480-16

- Dial Reading: 0-40-0
- Jeweled bearing
- Carbide contact point
- Two measuring directions
- Anti magnetic body
- Supplied with two clamps:
diameter Ø4mm and Ø8mm
- Optional accessory: styli



单位: mm

clamp (included)



Code	Range	Graduation	Accuracy	Hysteresis
2480-16*	1.6mm	0.01mm	18µm	3µm

* Supplied with manufacturer inspection certificate traceable to NIST USA